




Sizing Guide

Star Galaxy cell height



Item no. ESGXY...	Base (C x D) Width x Depth (cm)	Seat Width* (cm)	Seat Depth** (cm)
...1414-2F	36 x 36	32 - 36	33 - 37
...1416-2F	36 x 41	32 - 36	38 - 42
...1614-2F	41 x 36	37 - 41	33 - 37
...1616-2F	41 x 41	37 - 41	38 - 42
...1618-2F	41 x 46	37 - 41	43 - 47
...1816-2F	46 x 41	42 - 46	38 - 42
...1818-2F	46 x 46	42 - 46	43 - 47
...1820-2F	46 x 51	42 - 46	48 - 52
...2016-2F	51 x 41	47 - 51	38 - 42
...2018-2F	51 x 46	47 - 51	43 - 47
...2020-2F	51 x 51	47 - 51	48 - 52

...For the complete item number, see related product sheet.

Star Galaxy

Galaxy's large rear cells provide excellent support for users with gluteal muscle atrophy.

The base is approximately 2 cm wider on all sides (E), than the measurement from outer cell to outer cell (A and B)

The cushion base measurements (C x D) are indicated in all documentation.

- *Seat width:** If placed on a wheelchair the cushion base edges (E) may fold upwards. Hence the difference between the base measurements and recommended seat width.
- **Seat depth:** The seat depth is measured at the middle of the seat, from the front edge to the back upholstery. The measurement should be taken when the back upholstery is adjusted correctly. Place the cushion as far back on the seat as possible.